## **Calibration Certificate**

<u>Customer Information</u> <u>Equipment Information</u>

Customer:Lascar ElectronicsEquipment ID:L15-3733Address:4258 West 12th StModel No:EL-WIFI-TP+

Manufacturer: Lascar

**City, State Zip:** Erie PA 16505 **Serial No:** 98:8B:AD:00:74:3F

Contact:

**Phone:** (814) 835-0621

**Calibration Summary** 

 Calibrated:
 01/27/2016
 Temp:
 70F +/- 10
 As Found:
 In Tolerance

 Next Cal:
 01/27/2017
 Humidity:
 50% +/-20
 Result:
 Pass

Frequency: Yearly Technician: Angela Relyea

Remarks:

**Temperature Calibration (C)** 

ToleranceUnitsAs FoundResultProceduresTolerance (+/-)0.3 °CIn TolerancePassCSI-CWB-1 R4

| ,           |                | Limits       |       | As Found      |        |        | As Left (Cal Status) |              |               |
|-------------|----------------|--------------|-------|---------------|--------|--------|----------------------|--------------|---------------|
| <u>Desc</u> | <u>Nominal</u> | <u>Upper</u> | Lower | <u>Actual</u> | Error  | Result | <u>Actual</u>        | <u>Error</u> | <u>Result</u> |
|             | 25.00          | 25.30        | 24.70 | 24.95         | -00.05 | Pass   | 24.95                | -00.05       | Pass          |

## **Standard Equipment Used**

 Equip ID
 Model
 Description
 Cal Due

 Control Co.2
 4000
 12/08/2016

 Water Bath 1
 9006A11B
 03/05/2016

Control Solutions, Inc. certifies that the above equipment has been calibrated using instrumentation capable of producing results that are traceable through NIST to the International System of Units (SI). Control Solutions, Inc. is accredited to ISO/IEC 17025:2005.

Measurement uncertainty = 0.12 C, utilizing a coverage factor of K=2 to approximate 95% confidence level

This certificate cannot be reproduced, except in full, without the specific written permission of Control Solutions, Inc. 35851 Industrial Way, Suite D, St. Helens, OR 97051





www.lascarelectronics.com (814) 835 0621 4258 West 12th Street, Erie, PA 16505



Certificate: ID-524159